



Batch Number: **B 147**

Fluorescence analysis - Mean Fluorescence Intensity	
<i>Cryptosporidium</i> oocysts:	949.03
<i>Giardia</i> cysts:	1088.56
Non-specific binding assay	
Number of particles detected:	0
Control number of particles detected:	0

General

Safety: This product contains sodium azide. Handle with care.
Storage: 2-8°C. Do not freeze.
Catalogue Number: EST-CG80-D, EST-CG20-D, W-EST-CG80-D2, W-EST-CG20-D2
Preparation Date: 8 June 2021
Expiration Date: 9 June 2022

Quality Manager

Lucy Millican
10 June 2021

Manufactured By:

Biopoint Pty Ltd
Suite 16, 13A Narabang Way, Belrose, NSW, 2085
Tel. +61 2 8316 7939
www.biopoint.com.au
info@biopoint.com.au