

Certificate of Analysis

Batch Number: B 156

| Fluorescence analysis - Mean Fluorescence Intensity | | | | |
|---|--------|--|--|--|
| Cryptosporidium oocysts: | 979.69 | | | |
| Giardia cysts: | 924.54 | | | |
| Non-specific binding assay | | | | |
| Number of particles detected: | 0 | | | |
| Control number of particles detected: | 0 | | | |

General

| | Safety: | This product contains sodium azide. Handle with care. |
|--|-------------------|---|
| | Storage: | 2-8°C. Do not freeze. |
| | Catalogue Number: | EST-CG80-D, EST-CG20-D, JPN-EST-CG80-D, JPN-EST-CG20-D |
| | Preparation Date: | 23 August 2022 |
| | Expiration Date*: | 24 August 2023 |
| *Expiration Date applies to all kit contents | | |

Component Lots

| Mounting Medium Lot | MM- | 1002 / 1004 |
|---------------------|-----|-------------|
| Wash Buffer Lot | WB- | 1003 |
| Postive Control Lot | PC- | 1007 |
| DAPI Lot | D- | 1003 |

......

ocint

Quality Manager

1 min

Lucy Millican 30 August 2022

Manufactured By: Biopoint Pty Ltd Suite 16, 13A Narabang Way, Belrose, NSW, 2085 Tel. +61 2 8316 7939 www.biopoint.com.au info@biopoint.com.au

lssue: 6 Date:13/5/2022 Page 1 of 1