



Certificate of Analysis

Batch Number: B 158

Fluorescence analysis - Mean Fluorescence Intensity

Cryptosporidium oocysts: 1051.69

Giardia cysts: 918.21

Non-specific binding assay

Number of particles detected: 0

0

Control number of particles detected:

General

Safety: This product contains sodium azide.

Handle with care.

Storage: 2-8°C. Do not freeze.

Catalogue Number: EST-CG80-D, EST-CG20-D,

JPN-EST-CG80-D, JPN-EST-CG20-D

Preparation Date: 14 November 2022

Expiration Date*: 15 November 2023

*Expiration Date applies to all kit contents

Component Lots

Mounting Medium Lot MM- 1004

Wash Buffer Lot WB- 1003 Postive Control Lot PC- 1007

DAPI Lot D- 1003

Quality Manager

Lucy Millican 30 November 2022

Manufactured By:

Biopoint Pty Ltd Suite 16, 13A Narabang Way, Belrose, NSW, 2085

Tel. +61 2 8316 7939

www.biopoint.com.au

info@biopoint.com.au

Issue: 6 Date:13/5/2022 Page 1 of 1